



ACCURATE & EASY TO USE

Discover a wide selections of our
sound measurement equipments according to your
sound measurement needs

CA-01

SOUND CALIBRATOR



www.placidinstruments.com



PLACID INSTRUMENTS BV


Newtonlaan 115 Office No. 2.21 3584 BH UTRECHT, THE NETHERLANDS

Phone +31 302106021

email info@placidinstruments.com

website www.placidinstruments.com

PLACID Sound Calibrator CA-01



Calibrator for sound level meters, 1 kHz, levels 94 dB and 114 dB. Can be used for any sound level meter with a 1/2" microphone.

PLACID Sound level calibrator is used to calibrate sound level meters and other sound measurement equipment. You can calibrate 1" diameter microphones directly and 1/2" microphones using 1/2" adaptor supplied with the calibrator. With available adaptors, you can calibrate other microphones and instruments.



PLACID INSTRUMENTS BV

Newtonlaan 115 Office No. 2.21 3584 BH UTRECHT, THE NETHERLANDS

Phone +31 302106021

email info@placidinstruments.com

website www.placidinstruments.com

PLACID Sound Calibrator CA-01



SPECIFICATION

Sound Calibrator	class 1 according to EN/IEC 60942:2017 Class 1
Microphone Size	1/1" and 1/2"
Sound Pressure Level @ Reference conditions	94 ±0.25 dB / 114 ±0.25 dB
Frequency	1000 ±0.7% Hz
Total Harmonic Distortion + Noise	<2%
Power	one 9V battery 006P or IEC 6F22 or NEDA 1604.
Battery life	approx.40 hours.(Alkaline Battery)
Battery Test	Internal circuitry checks condition of battery continuously. Calibrator will not be operated when green LED turns to red LED, which means that battery voltage falls below acceptable range.
Dimensions	113(L)x63(W)x44(H)mm,4.4(L)x2.4(W)x1.7(H)inch
Weight	Approximately 170 gram (including battery)

REFERENCE CONDITIONS

Temperature	23°C (73°F)
Static press	103 hPa
Relative humidity	50%

FEATURES

- Sound calibrator with 2 levels, 94 dB and 114 dB at 1 kHz
- Accurate and very easy to use
- Can be used for calibration of 1/2" and 1/1" microphones
- Low battery indicator
- Complies with IEC60942: 2017 Class 1
- Complies with ANSI S1.40



Placid Instruments BV

Newtonlaan 115, Office No. 2.21

3584 BH UTRECHT

THE NETHERLANDS

Phone landline : +31 302106021

Email : info@placidinstruments.com

www.placidinstruments.com

ASK US!
HERE

